
Midmark M11 - v1.0.0
Total Cycles: 3305

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.7 F 0.0
2:00 189.0 F 0.9
4:00 212.5 F 6.0
6:00 230.7 F 11.5
8:00 229.5 F 7.3
10:00 241.6 F 13.0
12:00 240.5 F 11.0
14:00 238.7 F 9.7
16:00 253.1 F 16.8
18:00 265.2 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.7
2:00 271.6 F 28.4
3:00 272.5 F 28.7
4:00 271.5 F 28.1

Min 271.4 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 210.9 F 0.0
10:00 213.3 F 0.0
15:00 217.3 F 0.0
20:00 220.4 F 0.0
Dry Cycle Aborted

FILLING: 1:37
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 22:25
TOTAL CYCLE: 00:51:01

05/30/2023 01:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____