

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4060

Name: RITTER 1  
Sterilizer ID: V2178198  
11/27/2023 01:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.0 F 0.0  
2:00 121.2 F 1.5  
4:00 149.7 F 4.4  
6:00 175.8 F 8.0  
8:00 196.0 F 12.0  
10:00 216.1 F 4.6  
12:00 224.6 F 8.7  
14:00 238.9 F 13.0  
16:00 235.3 F 9.4  
18:00 240.0 F 10.3  
20:00 244.8 F 12.7  
22:00 256.0 F 18.6  
24:00 266.3 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.4 F 28.9  
2:00 272.5 F 29.4  
3:00 271.2 F 28.6  
4:00 272.3 F 29.3

Min 271.0 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 230.4 F 0.0  
10:00 230.0 F 0.0  
15:00 224.9 F 0.0  
20:00 222.0 F 0.0  
25:00 219.0 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 25:39  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:04:18  
CYCLE COMPLETE

11/27/2023 02:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_