
Midmark M11 - v1.0.0
Total Cycles: 1595

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2022 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.1 F 0.0
2:00 148.2 F 1.8
4:00 188.7 F 5.0
6:00 206.1 F 9.2
8:00 218.9 F 9.5
10:00 219.9 F 6.8
12:00 230.4 F 11.5
14:00 232.0 F 7.3
16:00 242.2 F 12.9
18:00 241.1 F 10.9
20:00 253.6 F 17.2
22:00 264.6 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.0 F 28.9
2:00 272.1 F 29.1
3:00 272.2 F 29.1
4:00 272.1 F 29.1

Min 271.5 F 28.5
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.9 F 0.0
5:00 224.6 F 0.0
10:00 235.6 F 0.0
15:00 234.8 F 0.0
20:00 229.6 F 0.0
25:00 226.9 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:23
CYCLE COMPLETE

02/03/2022 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____