

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4129

Name: RITTER2  
Sterilizer ID: V2313228  
11/18/2024 10:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 73.8 F 0.0  
4:00 85.7 F 0.0  
6:00 115.0 F 0.2  
8:00 158.5 F 0.5  
10:00 184.7 F 0.7  
12:00 200.6 F 2.0  
14:00 211.1 F 5.4  
16:00 222.1 F 8.9  
18:00 233.0 F 12.9  
20:00 233.5 F 9.0  
22:00 241.5 F 11.5  
24:00 243.5 F 12.4  
26:00 241.9 F 11.4  
28:00 254.5 F 18.0  
30:00 265.5 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.5 F 29.1  
2:00 272.2 F 29.2  
3:00 272.1 F 28.7  
4:00 272.8 F 29.5

Min 271.5 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
5:00 202.4 F 0.0  
10:00 212.6 F 0.0  
15:00 198.0 F 0.0  
20:00 195.2 F 0.0  
25:00 193.6 F 0.0  
30:00 206.2 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:09:35  
CYCLE COMPLETE

11/18/2024 12:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_