
Midmark M11 - v1.0.0
Total Cycles: 3895

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 168.1 F 3.5
4:00 184.0 F 7.8
6:00 212.6 F 13.6
8:00 222.6 F 7.7
10:00 235.8 F 13.0
12:00 236.9 F 10.3
14:00 235.3 F 8.1
16:00 249.8 F 15.4
18:00 262.0 F 22.3
20:00 272.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.3
2:00 271.7 F 28.5
3:00 272.3 F 29.3
4:00 272.4 F 29.0

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 216.0 F 0.0
10:00 230.9 F 0.0
15:00 227.5 F 0.0
20:00 226.4 F 0.0
25:00 232.7 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

10/12/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____