
Midmark M11 - v1.0.0
Total Cycles: 1600

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2022 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.9 F 0.0
2:00 157.0 F 3.1
4:00 171.5 F 6.9
6:00 199.6 F 11.8
8:00 215.2 F 5.5
10:00 228.7 F 10.5
12:00 232.3 F 7.3
14:00 242.8 F 12.8
16:00 241.7 F 11.2
18:00 254.8 F 17.9
20:00 266.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.3 F 29.0
2:00 272.3 F 28.7
3:00 271.6 F 28.5
4:00 272.8 F 29.2

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 226.0 F 0.0
10:00 235.2 F 0.0
15:00 227.5 F 0.0
20:00 227.9 F 0.0
25:00 230.6 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

02/07/2022 12:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____