
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001365
13:30 28/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:54
STERILIZING 4:00
136.0 C 220kPa 4:00
136.4 C 224kPa 4:30
136.4 C 222kPa 5:00
136.3 C 224kPa 5:30
136.3 C 224kPa 6:00
136.2 C 223kPa 6:30
136.3 C 223kPa 7:00
136.3 C 222kPa 7:30
136.3 C 222kPa 7:30
VENTING 7:30
AIR DRYING 8:22
CYCLE COMPLETE 23:22

Digital Signature #
830F871E8B925C05
