
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008356
9:02 26/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:08
136.5 C 223kPa 3:08
136.0 C 219kPa 3:38
135.8 C 218kPa 4:08
136.4 C 222kPa 4:38
136.4 C 222kPa 5:08
136.2 C 221kPa 5:38
136.1 C 221kPa 6:08
136.4 C 222kPa 6:38
136.4 C 222kPa 6:38
VENTING 6:38
AIR DRYING 7:27
CYCLE COMPLETE 35:39

Digital Signature #
B53622E3B25C7049
