
Midmark M11 - v1.0.0
Total Cycles: 3766

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2023 12:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.7 F 0.0
2:00 176.6 F 2.4
4:00 197.4 F 6.5
6:00 223.5 F 12.0
8:00 227.1 F 6.7
10:00 238.6 F 11.9
12:00 236.3 F 9.2
14:00 238.8 F 9.5
16:00 248.8 F 14.4
18:00 261.0 F 21.2
20:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.6
2:00 271.5 F 28.2
3:00 272.4 F 28.6
4:00 271.4 F 28.1

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 227.0 F 0.0
10:00 210.7 F 0.0
15:00 218.2 F 0.0
20:00 226.1 F 0.0
25:00 249.9 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

09/13/2023 01:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____