
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 003701
8:43 27/05/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:55
STERILIZING 3:58
136.0 C 220kPa 3:58
136.3 C 221kPa 4:28
136.1 C 221kPa 4:58
136.3 C 223kPa 5:28
136.4 C 224kPa 5:58
136.0 C 222kPa 6:28
136.3 C 221kPa 6:58
136.4 C 222kPa 7:28
136.4 C 222kPa 7:28
VENTING 7:28
AIR DRYING 8:15
CYCLE COMPLETE 8:16

Digital Signature #
93FE0B32CC17B5DE
