
Midmark M11 - v1.0.0
Total Cycles: 1525

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2022 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 194.8 F 0.3
4:00 214.1 F 4.2
6:00 230.6 F 9.3
8:00 236.3 F 9.5
10:00 242.6 F 12.1
12:00 240.9 F 10.7
14:00 238.9 F 9.8
16:00 253.8 F 17.1
18:00 266.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.6
2:00 271.5 F 28.2
3:00 272.5 F 28.7
4:00 271.4 F 28.0

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 237.4 F 0.0
10:00 217.5 F 0.0
15:00 214.3 F 0.0
20:00 217.9 F 0.0
25:00 226.4 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

01/13/2022 03:22 PM
ACCEPTED: X
REJECTED:
Operator: U02