
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001902
11:59 26/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 2:58
136.5 C 220kPa 2:58
136.4 C 226kPa 3:28
136.4 C 226kPa 3:58
136.4 C 225kPa 4:28
136.4 C 226kPa 4:58
136.4 C 225kPa 5:28
136.4 C 223kPa 5:58
136.3 C 225kPa 6:28
136.3 C 225kPa 6:28
VENTING 6:28
AIR DRYING 7:19
CYCLE COMPLETE 11:29

Digital Signature #
BE2032FFB304217D
