
Midmark M11 - v1.0.0
Total Cycles: 3563

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.7 F 0.0
2:00 149.8 F 0.0
4:00 159.3 F 1.6
6:00 196.3 F 3.9
8:00 200.4 F 7.6
10:00 213.5 F 11.6
12:00 224.1 F 5.8
14:00 234.9 F 10.4
16:00 234.0 F 8.4
18:00 244.4 F 12.9
20:00 242.3 F 11.4
22:00 254.0 F 17.3
24:00 264.6 F 23.6
26:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.1 F 28.8
2:00 271.2 F 28.1
3:00 272.1 F 29.0
4:00 271.6 F 28.5

Min 271.0 F 28.1
Max 272.3 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 210.3 F 0.0
10:00 216.4 F 0.0
15:00 212.7 F 0.0
20:00 230.1 F 0.0
25:00 231.8 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:04:30
CYCLE COMPLETE

07/21/2023 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____