
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002251
14:11 19/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 4:38
STERILIZING 5:25
136.4 C 223kPa 5:25
136.1 C 222kPa 5:55
136.1 C 223kPa 6:25
136.0 C 221kPa 6:55
136.4 C 222kPa 7:25
136.4 C 225kPa 7:55
136.5 C 226kPa 8:25
136.1 C 221kPa 8:55
136.3 C 225kPa 9:25
136.3 C 224kPa 9:55
136.0 C 220kPa 10:25
136.4 C 224kPa 10:55
136.4 C 225kPa 11:25
136.4 C 222kPa 11:55
136.4 C 223kPa 12:25
136.1 C 218kPa 12:55
136.4 C 218kPa 13:25
136.3 C 222kPa 13:55
136.1 C 221kPa 14:25
136.3 C 224kPa 14:55
136.4 C 225kPa 15:25
136.4 C 225kPa 15:25
VENTING 15:25
AIR DRYING 16:24
CYCLE COMPLETE 16:33

Digital Signature #

0177A5556C284C10
