
Midmark M11 - v1.0.5
Total Cycles: 2136

Name: RITTER2
Sterilizer ID: V2313228
03/09/2023 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.3 F 0.0
2:00 183.8 F 3.8
4:00 202.0 F 9.2
6:00 225.2 F 7.6
8:00 236.0 F 11.6
10:00 236.9 F 9.9
12:00 236.9 F 8.7
14:00 252.0 F 16.5
16:00 264.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.3 F 29.4
2:00 271.0 F 28.7
3:00 272.7 F 29.8
4:00 271.8 F 29.5

Min 271.0 F 28.7
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 231.9 F 0.0
10:00 230.1 F 0.0
15:00 229.2 F 0.0
20:00 226.9 F 0.0
25:00 228.7 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 17:48
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

03/09/2023 03:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____