

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001642  
14:23 30/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:14  
PRESSURIZING 2:19  
STERILIZING 3:20  
136.4 C 225kPa 3:20  
136.3 C 225kPa 3:50  
136.4 C 226kPa 4:20  
136.4 C 225kPa 4:50  
136.4 C 226kPa 5:20  
136.2 C 223kPa 5:50  
136.1 C 223kPa 6:20  
136.4 C 226kPa 6:50  
136.4 C 226kPa 6:50  
VENTING 6:50  
AIR DRYING 7:41  
CYCLE COMPLETE 7:53  
-----

Digital Signature #  
88ADDBFEF4943585  
-----