
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001845
9:15 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:36
STERILIZING 3:48
136.0 C 222kPa 3:48
136.3 C 225kPa 4:18
136.4 C 226kPa 4:48
136.4 C 225kPa 5:18
136.1 C 223kPa 5:48
136.3 C 225kPa 6:18
136.3 C 226kPa 6:48
136.2 C 225kPa 7:18
136.2 C 225kPa 7:18
VENTING 7:18
AIR DRYING 8:11
CYCLE COMPLETE 8:19

Digital Signature #
36CB1A670532D254
