

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000490  
14:08 09/04/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:21  
STERILIZING 4:35  
136.0 C 228kPa 4:35  
136.4 C 233kPa 5:05  
136.4 C 233kPa 5:35  
136.4 C 233kPa 6:05  
136.4 C 233kPa 6:35  
136.4 C 232kPa 7:05  
136.2 C 231kPa 7:35  
136.3 C 232kPa 8:05  
136.3 C 232kPa 8:05  
VENTING 8:05  
AIR DRYING 8:59  
CYCLE COMPLETE 14:19  
-----

Digital Signature #  
B56C1D38C8CB4B55  
-----