
Midmark M11 - v1.0.5
Total Cycles: 1060

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 179.0 F 2.0
4:00 189.6 F 6.4
6:00 208.3 F 11.8
8:00 225.5 F 7.6
10:00 239.0 F 13.2
12:00 239.8 F 11.2
14:00 238.8 F 10.1
16:00 252.3 F 16.9
18:00 263.8 F 23.7
20:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.2 F 29.8
2:00 272.1 F 29.3
3:00 271.0 F 28.9
4:00 271.7 F 29.7

Min 271.0 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 214.3 F 0.0
10:00 208.1 F 0.0
15:00 210.2 F 0.0
20:00 205.9 F 0.0
25:00 205.4 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

05/03/2022 09:03 AM
ACCEPTED: X
REJECTED:
Operator: U02