
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005954
9:09 12/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:58
STERILIZING 2:56
136.4 C 222kPa 2:56
136.2 C 217kPa 3:26
136.3 C 224kPa 3:56
136.4 C 224kPa 4:26
136.3 C 224kPa 4:56
136.4 C 224kPa 5:26
136.4 C 224kPa 5:56
136.3 C 218kPa 6:26
136.3 C 218kPa 6:26
VENTING 6:26
AIR DRYING 7:18
CYCLE COMPLETE 23:19

Digital Signature #
D33DAD0204C676A0
