

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1663

Name: RITTER2  
Sterilizer ID: V2313228  
10/26/2022 02:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.0 F 0.0  
2:00 189.0 F 4.0  
4:00 199.8 F 9.0  
6:00 222.8 F 7.6  
8:00 230.7 F 10.6  
10:00 233.7 F 8.3  
12:00 240.2 F 10.7  
14:00 246.3 F 14.1  
16:00 258.9 F 21.0  
18:00 270.0 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.2 F 29.6  
2:00 271.8 F 29.9  
3:00 271.9 F 29.4  
4:00 272.3 F 30.0  
  
Min 271.2 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 228.4 F 0.0  
10:00 244.1 F 0.0  
15:00 239.6 F 0.0  
20:00 228.0 F 0.0  
25:00 228.4 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 18:50  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:07  
CYCLE COMPLETE

10/26/2022 03:51 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_