
Midmark M11 - v1.0.0
Total Cycles: 4427

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2024 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.4 F 0.0
2:00 145.4 F 1.3
4:00 143.5 F 4.0
6:00 170.9 F 7.7
8:00 192.6 F 12.0
10:00 216.1 F 4.4
12:00 227.3 F 8.6
14:00 237.2 F 13.2
16:00 236.4 F 9.5
18:00 238.4 F 9.6
20:00 245.8 F 13.2
22:00 257.3 F 19.4
24:00 268.2 F 26.0
26:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.8 F 29.2
2:00 271.9 F 28.9
3:00 272.8 F 29.1
4:00 271.7 F 28.9

Min 271.5 F 28.6
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 224.7 F 0.0
10:00 205.8 F 0.0
15:00 206.7 F 0.0
20:00 195.6 F 0.0
25:00 197.3 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 26:01
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:04:30
CYCLE COMPLETE

04/04/2024 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____