
Midmark M11 - v1.0.5
Total Cycles: 3830

Name: RITTER2
Sterilizer ID: V2313228
08/02/2024 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.6 F 0.0
2:00 137.7 F 1.7
4:00 163.2 F 6.6
6:00 190.4 F 12.0
8:00 213.8 F 5.4
10:00 230.2 F 11.3
12:00 234.1 F 8.8
14:00 239.0 F 9.9
16:00 247.3 F 14.4
18:00 260.2 F 21.5
20:00 249.6 F 15.2
22:00 264.5 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 271.9 F 28.5
2:00 272.1 F 28.6
3:00 272.5 F 28.8
4:00 272.7 F 29.1

Min 271.8 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 208.3 F 0.0
10:00 208.6 F 0.0
15:00 218.6 F 0.0
20:00 210.5 F 0.0
25:00 209.6 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

08/02/2024 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____