

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4163

Name: RITTER 1  
Sterilizer ID: V2178198  
12/28/2023 10:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
2:00 175.9 F 2.1  
4:00 173.4 F 5.8  
6:00 195.0 F 10.4  
8:00 218.9 F 4.3  
10:00 232.8 F 9.6  
12:00 234.7 F 8.9  
14:00 242.1 F 12.0  
16:00 239.8 F 10.3  
18:00 253.0 F 16.7  
20:00 264.5 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.3 F 28.6  
2:00 271.6 F 28.5  
3:00 272.3 F 28.5  
4:00 271.5 F 28.4  
  
Min 271.2 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.2 F 0.0  
5:00 225.1 F 0.0  
10:00 208.4 F 0.0  
15:00 222.2 F 0.0  
20:00 211.8 F 0.0  
25:00 228.4 F 0.0  
30:00 221.3 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 21:57  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:00:30  
CYCLE COMPLETE

12/28/2023 11:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_