
Midmark M11 - v1.0.0
Total Cycles: 1766

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2022 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 167.2 F 0.6
4:00 164.7 F 3.1
6:00 186.5 F 6.7
8:00 206.6 F 11.5
10:00 220.4 F 5.5
12:00 231.7 F 10.7
14:00 231.7 F 7.1
16:00 245.8 F 13.4
18:00 243.7 F 11.9
20:00 256.7 F 18.6
22:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.2 F 28.8
2:00 271.9 F 28.3
3:00 271.2 F 28.0
4:00 272.2 F 28.9

Min 271.1 F 27.8
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 212.5 F 0.0
10:00 227.5 F 0.0
15:00 224.2 F 0.0
20:00 226.8 F 0.0
25:00 226.3 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

03/30/2022 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____