
Midmark M11 - v1.0.0
Total Cycles: 2025

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2022 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 188.5 F 3.7
4:00 202.1 F 8.0
6:00 230.9 F 13.5
8:00 230.1 F 8.4
10:00 240.1 F 11.7
12:00 240.8 F 11.4
14:00 238.5 F 9.8
16:00 252.9 F 16.7
18:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.1 F 28.5
2:00 272.4 F 29.0
3:00 271.5 F 28.1
4:00 272.4 F 28.9

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 236.9 F 0.0
10:00 228.1 F 0.0
15:00 225.3 F 0.0
20:00 228.1 F 0.0
25:00 226.1 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 1:24
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

06/15/2022 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____