

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3654

Name: RITTER2  
Sterilizer ID: V2313228  
05/28/2024 03:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.8 F 0.0  
2:00 165.5 F 3.1  
4:00 182.8 F 7.2  
6:00 204.8 F 12.3  
8:00 216.8 F 7.5  
10:00 238.9 F 12.6  
12:00 235.6 F 9.8  
14:00 236.3 F 9.0  
16:00 248.0 F 14.7  
18:00 259.8 F 21.3  
20:00 270.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.5  
1:00 272.8 F 30.1  
2:00 271.5 F 29.6  
3:00 273.1 F 29.8  
4:00 271.8 F 29.9  
  
Min 271.3 F 29.3  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 231.2 F 0.0  
10:00 222.1 F 0.0  
15:00 184.5 F 0.0  
20:00 181.2 F 0.0  
25:00 178.6 F 0.0  
30:00 177.4 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 21:20  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:00:40  
CYCLE COMPLETE

05/28/2024 04:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_