

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2499

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2023 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
2:00 184.5 F 3.6  
4:00 200.4 F 8.4  
6:00 224.1 F 9.9  
8:00 231.9 F 9.8  
10:00 233.5 F 8.5  
12:00 242.7 F 13.3  
14:00 242.8 F 12.3  
16:00 255.7 F 19.0  
18:00 266.9 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.6 F 29.9  
2:00 271.5 F 29.1  
3:00 272.0 F 29.7  
4:00 272.3 F 29.6  
  
Min 271.1 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 220.7 F 0.0  
10:00 233.2 F 0.0  
15:00 230.0 F 0.0  
20:00 229.2 F 0.0  
25:00 229.2 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

06/12/2023 11:35 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_