
Midmark M11 - v1.0.0
Total Cycles: 1069

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2021 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.2 F 0.0
2:00 182.2 F 1.5
4:00 178.1 F 4.6
6:00 204.0 F 8.8
8:00 222.6 F 9.0
10:00 226.6 F 8.3
12:00 238.6 F 13.1
14:00 237.6 F 10.5
16:00 238.1 F 8.3
18:00 248.4 F 14.8
20:00 260.2 F 21.3
22:00 270.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.1
2:00 271.6 F 28.8
3:00 272.5 F 29.1
4:00 271.5 F 28.7

Min 271.4 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 236.4 F 0.0
10:00 233.3 F 0.0
15:00 231.2 F 0.0
20:00 229.0 F 0.0
25:00 193.3 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:43
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

08/04/2021 09:05 AM
ACCEPTED: X
REJECTED:
Operator: U01