

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000142  
7:22 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:15  
PRESSURIZING 3:23  
STERILIZING 4:45  
136.0 C 220kPa 4:45  
136.4 C 225kPa 5:15  
136.4 C 225kPa 5:45  
136.4 C 223kPa 6:15  
136.4 C 222kPa 6:45  
136.4 C 224kPa 7:15  
136.4 C 224kPa 7:45  
136.3 C 223kPa 8:15  
136.3 C 223kPa 8:15  
VENTING 8:15  
AIR DRYING 9:07  
CYCLE COMPLETE 9:34  
-----

Digital Signature #  
87CB6AA373B9FD85  
-----