
Midmark M11 - v1.0.0
Total Cycles: 1222

Name: RITTER 1
Sterilizer ID: V2178198
09/30/2021 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.4 F 0.0
2:00 146.4 F 1.3
4:00 169.8 F 4.3
6:00 194.9 F 8.9
8:00 212.0 F 6.0
10:00 225.0 F 8.2
12:00 237.5 F 13.3
14:00 237.8 F 10.1
16:00 236.1 F 8.2
18:00 249.5 F 14.9
20:00 261.6 F 21.8
22:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.8 F 28.9
2:00 271.6 F 28.2
3:00 272.9 F 29.1
4:00 272.0 F 28.4

Min 271.5 F 28.2
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 214.5 F 0.0
10:00 216.2 F 0.0
15:00 213.2 F 0.0
20:00 213.0 F 0.0
25:00 211.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

09/30/2021 10:56 AM
ACCEPTED: X
REJECTED:
Operator: U02