
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001580
9:57 09/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:41
STERILIZING 3:57
136.4 C 222kPa 3:57
136.0 C 221kPa 4:27
136.1 C 221kPa 4:57
136.4 C 225kPa 5:27
136.4 C 223kPa 5:57
136.4 C 224kPa 6:27
136.2 C 222kPa 6:57
136.3 C 223kPa 7:27
136.3 C 223kPa 7:27
VENTING 7:27
AIR DRYING 8:18
CYCLE COMPLETE 17:43

Digital Signature #
37D1C13FB4620AFB
