
Midmark M11 - v1.0.5
Total Cycles: 197

Name: RITTER2
Sterilizer ID: V2313228
07/13/2021 06:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 73.8 F 0.0
4:00 75.9 F 0.1
6:00 94.1 F 1.8
8:00 128.1 F 4.4
10:00 154.5 F 7.6
12:00 173.4 F 11.2
14:00 213.9 F 4.7
16:00 206.5 F 7.7
18:00 219.7 F 11.4
20:00 228.8 F 7.3
22:00 238.4 F 12.3
24:00 237.7 F 10.2
26:00 249.1 F 15.7
28:00 259.4 F 21.3
30:00 267.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.1
1:00 272.0 F 30.0
2:00 271.7 F 30.0
3:00 272.2 F 30.2
4:00 271.7 F 30.4

Min 271.2 F 30.0
Max 272.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 214.2 F 0.0
10:00 208.4 F 0.0
15:00 199.2 F 0.0
20:00 197.6 F 0.0
25:00 200.0 F 0.0
30:00 196.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:49
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:52
CYCLE COMPLETE

07/13/2021 08:09 AM
ACCEPTED: X
REJECTED:
Operator: U01