

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2784

Name: RITTER2  
Sterilizer ID: V2313228  
08/17/2023 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.8 F 0.0  
2:00 134.0 F 0.0  
4:00 159.8 F 0.0  
6:00 194.8 F 0.0  
8:00 212.2 F 2.3  
10:00 224.5 F 6.4  
12:00 235.1 F 10.8  
14:00 234.3 F 8.3  
16:00 243.3 F 12.1  
18:00 245.1 F 13.1  
20:00 244.8 F 12.8  
22:00 257.7 F 19.7  
24:00 268.9 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.6 F 29.1  
2:00 272.1 F 29.2  
3:00 272.2 F 28.8  
4:00 272.2 F 29.3

Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
5:00 206.5 F 0.0  
10:00 208.6 F 0.0  
15:00 210.1 F 0.0  
20:00 211.0 F 0.0  
25:00 211.7 F 0.0  
30:00 210.7 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:04  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:03:23  
CYCLE COMPLETE

08/17/2023 01:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_