
Midmark M11 - v1.0.5
Total Cycles: 29

Name: RITTER2
Sterilizer ID: V2313228
05/19/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 179.0 F 2.5
4:00 191.8 F 6.8
6:00 212.0 F 12.5
8:00 227.7 F 8.0
10:00 237.8 F 11.9
12:00 241.6 F 12.0
14:00 241.5 F 11.3
16:00 254.9 F 18.2
18:00 266.6 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.0
2:00 272.4 F 29.5
3:00 272.0 F 28.7
4:00 272.7 F 29.5

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 227.3 F 0.0
10:00 227.0 F 0.0
15:00 207.3 F 0.0
20:00 218.6 F 0.0
25:00 218.6 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

05/19/2021 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____