
Midmark M11 - v1.0.0
Total Cycles: 2067

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2022 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.2 F 0.0
2:00 182.2 F 2.2
4:00 190.8 F 6.8
6:00 212.4 F 12.4
8:00 226.7 F 6.6
10:00 237.2 F 11.7
12:00 234.9 F 8.5
14:00 241.4 F 10.6
16:00 246.1 F 13.1
18:00 258.4 F 19.7
20:00 269.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.9
2:00 272.0 F 28.4
3:00 271.1 F 28.0
4:00 272.3 F 29.0

Min 271.1 F 27.9
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 222.9 F 0.0
10:00 227.2 F 0.0
15:00 231.5 F 0.0
20:00 217.8 F 0.0
25:00 221.3 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

06/28/2022 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____