
Midmark M11 - v1.0.5
Total Cycles: 758

Name: RITTER2
Sterilizer ID: V2313228
02/03/2022 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.0 F 0.0
2:00 192.6 F 3.6
4:00 192.3 F 8.3
6:00 213.8 F 10.3
8:00 227.0 F 9.3
10:00 233.6 F 9.2
12:00 243.1 F 13.4
14:00 243.8 F 12.6
16:00 256.9 F 19.6
18:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.8
2:00 272.4 F 29.4
3:00 271.4 F 28.8
4:00 271.5 F 29.3

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 228.8 F 0.0
10:00 235.6 F 0.0
15:00 227.6 F 0.0
20:00 228.7 F 0.0
25:00 229.2 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

02/03/2022 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____