
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000099
10:20 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:52
STERILIZING 5:20
136.4 C 227kPa 5:20
136.3 C 225kPa 5:50
136.4 C 226kPa 6:20
136.4 C 227kPa 6:50
136.4 C 226kPa 7:20
136.3 C 225kPa 7:50
136.4 C 225kPa 8:20
136.4 C 227kPa 8:50
136.4 C 227kPa 8:50
VENTING 8:50
AIR DRYING 9:58
CYCLE COMPLETE 22:17

Digital Signature #
F596D6FE16C33BC9
