

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1321

Name: RITTER 1  
Sterilizer ID: V2178198  
10/29/2021 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
2:00 189.0 F 3.2  
4:00 201.2 F 8.5  
6:00 218.9 F 9.7  
8:00 230.2 F 9.0  
10:00 237.4 F 10.3  
12:00 242.3 F 12.1  
14:00 240.9 F 10.7  
16:00 254.4 F 17.5  
18:00 266.1 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 271.7 F 28.2  
2:00 271.9 F 28.2  
3:00 272.3 F 28.6  
4:00 272.7 F 28.9  
  
Min 271.6 F 28.2  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 243.0 F 0.0  
10:00 226.1 F 0.0  
15:00 227.7 F 0.0  
20:00 222.9 F 0.0  
25:00 221.5 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 2:26  
HEATING: 19:30  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

10/29/2021 12:21 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02