
Midmark M11 - v1.0.5
Total Cycles: 3009

Name: RITTER2
Sterilizer ID: V2313228
10/13/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.6 F 0.0
2:00 140.4 F 1.9
4:00 163.2 F 5.5
6:00 187.4 F 9.9
8:00 212.1 F 5.7
10:00 222.7 F 8.4
12:00 235.9 F 13.5
14:00 237.5 F 11.0
16:00 236.7 F 9.4
18:00 250.7 F 16.2
20:00 262.5 F 23.1
22:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.5 F 29.7
4:00 271.6 F 29.0

Min 271.4 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 203.8 F 0.0
10:00 204.7 F 0.0
15:00 202.4 F 0.0
20:00 199.4 F 0.0
25:00 199.5 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:49
CYCLE COMPLETE

10/13/2023 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____