
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005937
11:47 10/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:13
STERILIZING 3:12
136.5 C 220kPa 3:12
136.4 C 224kPa 3:42
136.4 C 224kPa 4:12
136.4 C 225kPa 4:42
136.4 C 223kPa 5:12
136.1 C 223kPa 5:42
136.1 C 218kPa 6:12
136.3 C 222kPa 6:42
136.3 C 222kPa 6:42
VENTING 6:42
AIR DRYING 7:31
CYCLE COMPLETE 7:33

Digital Signature #
CA7AC5598EE83063
