
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000689
10:06 15/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:11
STERILIZING 3:12
136.0 C 230kPa 3:12
135.9 C 229kPa 3:42
136.4 C 233kPa 4:12
136.2 C 232kPa 4:42
136.4 C 233kPa 5:12
136.4 C 233kPa 5:42
136.2 C 230kPa 6:12
136.4 C 231kPa 6:42
136.4 C 231kPa 6:42
VENTING 6:42
AIR DRYING 7:35
CYCLE COMPLETE 7:37

Digital Signature #
F468CD50685253F8
