
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001684
16:30 11/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:31
STERILIZING 3:31
136.0 C 220kPa 3:31
136.4 C 225kPa 4:01
136.4 C 225kPa 4:31
136.3 C 223kPa 5:01
136.4 C 224kPa 5:31
136.1 C 221kPa 6:01
136.2 C 224kPa 6:31
136.1 C 222kPa 7:01
136.1 C 222kPa 7:01
VENTING 7:01
AIR DRYING 7:52
CYCLE COMPLETE 8:22

Digital Signature #
B4B063FA40E8E1CC
