
Midmark M11 - v1.0.0
Total Cycles: 5096

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2024 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.4 F 0.0
4:00 88.2 F 0.0
6:00 121.7 F 0.0
8:00 152.9 F 0.2
10:00 180.5 F 0.9
12:00 197.9 F 3.6
14:00 207.3 F 6.8
16:00 219.9 F 10.2
18:00 230.7 F 8.4
20:00 234.4 F 9.1
22:00 242.9 F 12.1
24:00 242.0 F 11.3
26:00 238.4 F 9.6
28:00 252.3 F 16.2
30:00 263.8 F 22.9
32:00 271.6 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.5
2:00 272.3 F 28.7
3:00 271.9 F 28.1
4:00 272.7 F 28.9

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 204.5 F 0.0
10:00 195.6 F 0.0
15:00 250.8 F 0.0
20:00 231.3 F 0.0
25:00 232.2 F 0.0
30:00 221.4 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:03
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:14
CYCLE COMPLETE

11/15/2024 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____