
Midmark M11 - v1.0.0
Total Cycles: 3142

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.0 F 0.0
2:00 183.5 F 0.0
4:00 195.7 F 2.7
6:00 214.6 F 7.9
8:00 233.3 F 13.2
10:00 233.3 F 9.2
12:00 240.2 F 9.4
14:00 245.0 F 13.0
16:00 243.6 F 11.9
18:00 256.8 F 18.9
20:00 268.5 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.7
2:00 271.4 F 28.2
3:00 272.6 F 28.8
4:00 271.7 F 28.1

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 212.9 F 0.0
10:00 220.9 F 0.0
15:00 230.2 F 0.0
20:00 228.6 F 0.0
25:00 229.7 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

04/21/2023 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____