

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1982

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2022 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.5 F 0.0  
2:00 160.0 F 2.8  
4:00 168.9 F 6.8  
6:00 208.6 F 11.9  
8:00 221.9 F 6.3  
10:00 233.9 F 11.3  
12:00 231.9 F 7.3  
14:00 244.1 F 13.2  
16:00 242.4 F 11.3  
18:00 255.1 F 18.0  
20:00 266.4 F 24.8  
22:00 273.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.8 F 29.0  
2:00 271.4 F 28.8  
3:00 272.7 F 28.9  
4:00 271.5 F 28.9

Min 271.1 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
5:00 236.4 F 0.0  
10:00 221.8 F 0.0  
15:00 221.5 F 0.0  
20:00 196.6 F 0.0  
25:00 209.4 F 0.0  
30:00 220.5 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 22:19  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:44  
CYCLE COMPLETE

05/31/2022 03:39 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02