
Midmark M11 - v1.0.0
Total Cycles: 4971

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2024 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 149.7 F 0.5
4:00 164.6 F 2.0
6:00 194.5 F 6.2
8:00 217.4 F 11.1
10:00 225.5 F 5.6
12:00 234.0 F 10.7
14:00 233.7 F 7.4
16:00 244.3 F 13.4
18:00 243.5 F 12.0
20:00 255.9 F 18.5
22:00 267.3 F 25.4
24:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.5 F 28.6
2:00 271.7 F 28.6
3:00 271.8 F 28.7
4:00 272.1 F 28.9

Min 271.3 F 28.6
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 244.5 F 0.0
10:00 203.3 F 0.0
15:00 239.3 F 0.0
20:00 223.6 F 0.0
25:00 224.4 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

10/07/2024 12:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____