

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.4uS / 1.5ppm  
CYCLE NUMBER 001868  
8:27 18/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 2:03  
STERILIZING 3:13  
136.4 C 225kPa 3:13  
136.4 C 226kPa 3:43  
136.4 C 226kPa 4:13  
136.4 C 226kPa 4:43  
136.3 C 223kPa 5:13  
136.3 C 224kPa 5:43  
136.4 C 226kPa 6:13  
136.4 C 226kPa 6:43  
136.4 C 226kPa 6:43  
VENTING 6:43  
AIR DRYING 7:34  
CYCLE COMPLETE 7:50

-----  
Digital Signature #  
306F972BD6445998  
-----