

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1667

Name: RITTER 1  
Sterilizer ID: V2178198  
02/24/2022 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.2 F 0.0  
2:00 102.4 F 0.0  
4:00 110.8 F 0.9  
6:00 141.9 F 2.3  
8:00 170.0 F 4.6  
10:00 189.5 F 7.9  
12:00 206.5 F 11.6  
14:00 215.2 F 4.1  
16:00 226.2 F 8.3  
18:00 237.1 F 12.6  
20:00 236.3 F 9.3  
22:00 240.7 F 9.8  
24:00 246.0 F 13.1  
26:00 257.5 F 19.3  
28:00 268.1 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.7 F 28.9  
2:00 271.7 F 28.1  
3:00 272.6 F 29.0  
4:00 272.2 F 28.4

Min 271.6 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 202.8 F 0.0  
10:00 213.6 F 0.0  
15:00 217.3 F 0.0  
20:00 213.0 F 0.0  
25:00 218.2 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 29:22  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:07:30  
CYCLE COMPLETE

02/24/2022 11:31 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02