
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002091
9:40 24/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:11
STERILIZING 4:21
136.0 C 220kPa 4:21
136.4 C 225kPa 4:51
136.4 C 224kPa 5:21
136.2 C 223kPa 5:51
136.4 C 226kPa 6:21
136.4 C 225kPa 6:51
136.2 C 223kPa 7:21
136.4 C 226kPa 7:51
136.4 C 226kPa 7:51
VENTING 7:51
AIR DRYING 8:44
CYCLE COMPLETE 8:53

Digital Signature #
4120913F58C04D58
